

상세규격서

구분	품명	단위	수량	비고
1	Ø1" Diaphragm Shutters with Controller (MODEL : SHB1T)	ea	6	
2	Angle Post Clamps (MODEL : RA45/M-P5)	ea	10	
3	Angle Post Clamps (MODEL : RA90M-P5)	ea	20	
4	Angle Post Clamps (MODEL : RA180M-P5)	ea	20	
5	Angle Post Clamps (MODEL : RA360/M)	ea	30	
6	Rotating Clamp (MODEL : SWC/M-5)	ea	10	
7	Mounting Joint (MODEL : PJ301/M)	ea	20	
8	Mounting Joint (MODEL : PJ302/M)	ea	20	

상세내역

1. 구성

1. Ø1" Diaphragm Shutters with Controller (MODEL : SHB1T)
2. Angle Post Clamps (MODEL : RA45/M-P5)
3. Angle Post Clamps (MODEL : RA90M-P5)
4. Angle Post Clamps (MODEL : RA180M-P5)
5. Angle Post Clamps (MODEL : RA360/M)
6. Rotating Clamp (MODEL : SWC/M-5)
7. Mounting Joint (MODEL : PJ301/M)
8. Mounting Joint (MODEL : PJ302/M)

2. 구성

1. Ø1" Diaphragm Shutters with Controller (MODEL : SHB1T)



- Operating Frequency : 0 - 15 Hz
- Minimum Exposure Pulse : 28 ms

- Open Time (Typical) : 10 ms
- Close Time (Typical) : 9 ms

2. Angle Post Clamps (MODEL : RA45/M-P5)



- Right-Handed 45° Angle Clamp for $\varnothing 1/2$ " Posts
- 5 mm Hex
- 5 Pack

3. Angle Post Clamps (MODEL : RA90M-P5)



- Right-Angle Clamp for $\varnothing 1/2$ " Posts
- 5 mm Hex
- 5 Pack

4. Angle Post Clamps (MODEL : RA180M-P5)



- Right-Angle End Clamp for $\text{\O}1/2''$ Posts
- M6 Stud and 5 mm Hex
- 5 Pack

5. Angle Post Clamps (MODEL : RA360/M)



- Parallel Clamp for $\text{\O}1/2''$ Posts
- M4 Counterbore and 5 mm Hex

6. Rotating Clamp (MODEL : SWC/M-5)



- Rotating Clamp for Ø1/2" Posts
- 360° Continuously Adjustable
- 5 mm Hex
- 5 Pack

7. Mounting Joint (MODEL : PJ301/M)



- Centered Mounting Post Joist
- M4 Mounting Hardware

8. Mounting Joint (MODEL : PJ302/M)



- Offset Mounting Post Joist
- M4 Mounting Hardware

3. 기타사항

가. 설치위치 : 경북 구미시 구미대로 350-27 XR 디바이스 개발 지원센터 106호

나. 설치개요 : 규격서의 조건에 맞는 광학 물품 납품

다. 납품기한 : 계약일로 부터 56 이내

라. 계약의 이행

- 1) 본 규격서에 대한 주계약자(이하 “계약자” 라 한다)와 발주기관 간의 해석상 이견은 발주기관의 해석에 따라야 하며, 본 규격서에 특별히 명시되지 않은 사항은 정부계약법 등의 규정과 일반 관례에 따른다.
- 2) 발주기관은 계약내용 전부에 대한 권리를 행사할 수 있으며 계약자는 이에 대하여 전적으로 동의해야 한다.
- 3) 본 규격서에 명시된 모든 조항은 본 사업의 원활한 추진에 필요한 최소한의 사항만을 규정하였으므로 상세히 기술하지 않았거나 누락된 사항에 대해서도 정상운영 및 서비스에 문제가 발생되지 않도록 계약자는 사전 조치를 취하여야 한다.
- 4) 본 규격서에서 요구하는 성능 및 특성에 대해 계약자의 오인 또는 누락된 상태로 계약 체결되어, 계약 후 이를 보완하기 위한 설계 변경, 물품 등을 추가해야 할 경우 계약자 부담으로 시행한다.
- 5) 계약자는 공급하는 제품의 성능, 품질 등에 대한 최종적인 책임을 진다.

4. 기타조건

가. 납품기한 : 계약 후 56 이내

다. 규격관련 문의사항 : 강필성 선임 : 010-4261-0006